

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Luna® SARS-CoV-2 RT-qPCR Multiplex Assay Kit

Catalog Number: E3019S
Packaging Lot Number: 10123078
Expiration Date: 09/2022
Storage Temperature: -20°C

Specification Version: PS-E3019S/L v1.0

Luna® SARS-CoV-2 RT-qPCR Multiplex Assay Kit Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
S0631AVIAL	SARS-CoV-2 Primer/Probe Mix (N1/N2/RP)	10083037	Pass	
N2117SVIAL	SARS-CoV-2 Positive Control (N gene)	10100711	Pass	
M3019SVIAL	Luna® Probe One-Step RT-qPCR 4X Mix with UDG	10121264	Pass	
B1502AVIAL	Nuclease-free Water	10118043	Pass	

Assay Name/Specification	Lot # 10123078
Functional Testing (RT-qPCR, SARS-CoV-2)	Pass
Luna® SARS-CoV-2 RT-qPCR Multiplex Assay Kit is functionally tested and compared to a control lot using 10 ng of human total RNA spiked with synthetic SARS-CoV-2 RNA in	
an RT-qPCR assay, resulting in a standard curve with a calculated qPCR efficiency of	
90-110%, and a dynamic range of 5 orders of magnitude.	
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in Luna®	1 4.50
SARS-CoV-2 RT-qPCR Multiplex Assay Kit and meet the designated specifications.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E3019S / Lot: 10123078

Page 1 of 2

Christie Vazquez
Production Scientist
06 Oct 2021

Michael Tonello

Packaging Quality Control Inspector

06 Oct 2021

E3019S / Lot: 10123078

Page 2 of 2