

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> Ultra<sup>™</sup> DNA Library Prep Kit for Illumina<sup>®</sup>  
 Catalog Number: E7370L  
 Lot Number: 10021060  
 Expiration Date: 07/2019  
 Storage Temperature: -20°C  
 Specification Version: PS-E7370S/L v1.0

NEBNext <sup>®</sup> Ultra <sup>™</sup> DNA Library Prep Kit for Illumina <sup>®</sup> Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7374AAVIAL	NEBNext <sup>®</sup> Ligation Enhancer	0031802	Pass
E7373AAVIAL	Blunt/TA Ligase Master Mix	0371802	Pass
E7372AAVIAL	NEBNext <sup>®</sup> End Repair Reaction Buffer	0211802	Pass
E7371AAVIAL	NEBNext <sup>®</sup> End Prep Enzyme Mix	0151802	Pass
E6625AAVIAL	NEBNext <sup>®</sup> Q5 <sup>®</sup> Hot Start HiFi PCR Master Mix	0091802	Pass

Assay Name/Specification	Lot # 10021060
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> Ultra<sup>™</sup> DNA Library Prep Kit for Illumina<sup>®</sup> and meet the designated specifications.</p> <p><b>Functional Testing (Library Construction, DNA)</b>            Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input DNA amounts are sequenced together on the same Illumina<sup>®</sup> flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p>	<p>Pass</p> <p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
28 Aug 2018

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
28 Aug 2018