

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat)

Catalog Number: E7750L
Packaging Lot Number: 10117851
Expiration Date: 11/2022
Storage Temperature: -20°C

Specification Version: PS-E7750S/L/X v1.0

NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E7753-3VIAL	NEBNext® DNase I	10089808	Pass	
E7752-3VIAL	NEBNext® Thermostable RNase H	10089809	Pass	
E7751-3VIAL	NEBNext® Globin & rRNA Depletion Solutio	10117646	Pass	
E6317-3VIAL	Nuclease-free Water	10089812	Pass	
E6315-3VIAL	DNase I Reaction Buffer	10089811	Pass	
E6314-3VIAL	NEBNext® Probe Hybridization Buffer	10089813	Pass	
E6312-3VIAL	RNase H Reaction Buffer	10089810	Pass	

Assay Name/Specification	Lot # 10117851
Functional Testing (Globin and rRNA Depletion) The NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) is functionally validated using commercially available human RNA extracted from blood. After treatment with	Pass
the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA and less than 5% globin RNA.	
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7750L / Lot: 10117851 Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Christine Sumner Production Scientist 29 Jul 2021

Josh Hersey Packaging Quality Control Inspector

29 Jul 2021



E7750L / Lot: 10117851

Page 2 of 2