

New England Biolabs Certificate of Analysis

Product Name: Amylose Resin High Flow
 Catalog Number: E8022S
 Packaging Lot Number: 10160897
 Expiration Date: 09/2024
 Storage Temperature: 4°C
 Specification Version: PS-E8022S/L v2.0

| Amylose Resin High Flow Component List | | | |
|--|-------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| E8022SVIAL | Amylose Resin High Flow | 10114739 | Pass |

| Assay Name/Specification | Lot # 10160897 |
|--|----------------|
| <p>Functional Binding Assay (Resin Binding Capacity) Amylose Resin High Flow (1 ml) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosinΔSal fusion protein (8 ml) was then passed through the column at 25°C, then washed with column buffer and the target protein eluted with ≥4 ml of column buffer containing 10 mM maltose. Binding capacity was determined to be >4 mg MBP5*-paramyosinΔSal /ml of resin based on A280 of the eluate.</p> | Pass |
| <p>Functional Binding Assay (Resin Binding Specificity) Amylose Resin High Flow (1 ml) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosinΔSal fusion protein (8 ml) was then passed through the column at 25°C, and then washed with column buffer. The target protein was eluted with ≥4 ml of column buffer containing 10 mM maltose. SDS-PAGE of the eluate on a 10-20% Tris-Glycine gel confirms low non-specific binding of extract proteins and high isolation of target.</p> | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Brad Landgraf
Production Scientist
12 Aug 2022



Michael Tonello
Packaging Quality Control Inspector
12 Aug 2022