

be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	Quick-Load® Purple 100 bp DNA Ladder
Catalog Number:	N0551S
Concentration:	50 μg/ml
Unit Definition:	N/A
Packaging Lot Number:	10129477
Expiration Date:	07/2023
Storage Temperature:	4°C
Storage Conditions:	2.5 % Ficoll 400 , 10 mM EDTA , 3.3 mM Tris-HCI (pH 8.0), 0.001 %
	Dye 2 , 0.02 % Dye 1
Specification Version:	PS-N0551S v1.0

Quick-Load® Purple 100 bp DNA Ladder Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N0551SVIAL	Quick-Load® Purple 100 bp DNA Ladder	10115679	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10121372	Pass	

Assay Name/Specification	Lot # 10129477
<b>Electrophoretic Pattern (Marker)</b> The banding pattern of Quick-Load <sup>®</sup> Purple 100 bp DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 $\mu$ g/ml.	Pass
<b>DNA Concentration (A260)</b> The concentration of Quick-Load® Purple 100 bp DNA Ladder is between 50 and 55 μg/ml as determined by UV absorption at 260 nm.	Pass
<b>Non-Specific DNase Activity (DNA, 16 hour)</b> A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of Quick-Load® Purple 100 bp DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Nulhin

Vanessa Mathieu-Sheltry Production Scientist 18 Nov 2021

Michae 011.

Michael Tonello Packaging Quality Control Inspector 18 Nov 2021

