

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: 1 kb Plus DNA Ladder

Catalog Number:N3200LConcentration:1,000 μg/mlLot Number:10010859Expiration Date:06/2020Storage Temperature:-20°C

Specification Version: PS-N3200S/L v2.0

1 kb Plus DNA Ladder Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N3200LVIAL	1 kb Plus DNA Ladder	10010737	Pass
B7025SVIAL	Gel Loading Dve. Purple (6X), no SDS	0261803	Pass

Assay Name/Specification	Lot # 10010859
A260/A280 Assay	Pass
The ratio of UV absorption of 1 kb Plus DNA Ladder at 260 and 280 nm is between 1.8 and 2.0.	
DNA Concentration (A260)	Pass
The concentration of 1 kb Plus DNA Ladder is between 1000 and 1050 µg/ml as	
determined by UV absorption at 260 nm.	
Electrophoretic Pattern (Marker)	Pass
The banding pattern of 1 kb Plus DNA Ladder on a 1.2% agarose gel shows discrete,	
clearly identifiable bands at each band of the marker, when stained with Ethidium	
Bromide at a concentration of 0.5 μg/ml.	
Non-Specific DNase Activity (DNA, 16 hour)	Pass
A 50 µl reaction in 1X NEBuffer 2 containing 5 µg of 1 kb Plus DNA Ladder incubated	
for 16 hours at 37°C results in a DNA pattern free of detectable nuclease	
degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



N3200L / Lot: 10010859

Page 1 of 2

Vanessa Mathieu-Sheltry
Production Scientist

31 May 2018

Michael Tonello

Packaging Quality Control Inspector

04 Jun 2018