

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

2-Log DNA Ladder (0.1-10.0 kb)
N3200S/L
1 mg/ml
N/A
0861612
12/2016
12/2018
-20°C
10 mM Tris-HCl (pH 8.0), 1 mM EDTA
PS-N3200S/L v1.0
15 Dec 2016

Assay Name/Specification (minimum release criteria)	Lot #0861612
A260/A280 Assay - The ratio of UV absorption of 2-Log DNA Ladder (0.1-10.0 kb) at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) - The concentration of 2-Log DNA Ladder (0.1-10.0 kb) is between 1000 and 1050 μ g/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Marker) - The banding pattern of 2-Log DNA Ladder (0.1-10.0 kb) on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 μ g/ml.	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 µl reaction in 1X NEBuffer 2 containing 5 µg of 2-Log DNA Ladder (0.1-10.0 kb) incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

otunen

Authorized by Derek Robinson 15 Dec 2016



Mulhiceheter

Inspected by Vanessa Mathieu-Sheltry 20 Dec 2016