

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: 1 kb Plus DNA Ladder

Catalog Number: N3200S
Concentration: 1,000 µg/ml
Packaging Lot Number: 10122953
Expiration Date: 08/2023
Storage Temperature: -20°C

Specification Version: PS-N3200S/L v2.0

1 kb Plus DNA Ladder Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N3200SVIAL	1 kb Plus DNA Ladder	10118553	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10121372	Pass	

Assay Name/Specification	Lot # 10122953
Non-Specific DNase Activity (DNA, 16 hour) A 50 μl reaction in 1X NEBuffer 2 containing 5 μg of 1 kb Plus DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease	Pass
degradation as determined by agarose gel electrophoresis.	Dana
A260/A280 Assay The ratio of UV absorption of 1 kb Plus DNA Ladder at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) The concentration of 1 kb Plus DNA Ladder is between 1000 and 1050 μg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Marker) The banding pattern of 1 kb Plus DNA Ladder on a 1.2% agarose gel shows discrete,	Pass
clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N3200S / Lot: 10122953

Page 1 of 2

Vanessa Mathieu-Sheltry

Production Scientist 29 Sep 2021 Michael Tonello
Packaging Quality Control Inspector

29 Sep 2021