

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	1 kb DNA Ladder
Catalog #:	N3232S/L
Concentration:	500 µg/ml
Unit Definition:	N/A
<i>Lot #</i> :	1451711
Assay Date:	11/2017
Expiration Date:	11/2019
Storage Temp:	-20°C
Storage Conditions:	10 mM Tris-HCl (pH 8.0), 1 mM EDTA
Specification Version:	PS-N3232S/L v1.0
Effective Date:	25 Sep 2017

Assay Name/Specification (minimum release criteria)	Lot #1451711
A260/A280 Assay - The ratio of UV absorption of 1 kb DNA Ladder at 260 and 280 nm is between 1.8 and 2.0.	Pass
<b>DNA Concentration (A260)</b> - The concentration of 1 kb DNA Ladder is between 500 and 550 $\mu$ g/ml as determined by UV absorption at 260 nm.	Pass
<b>Electrophoretic Pattern</b> - The banding pattern of 1 kb DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each size fragment of the marker, when stained with Ethidium Bromide at a concentration of 0.5 $\mu$ g/ml.	Pass
<b>Non-Specific DNase Activity (DNA, 16 hour)</b> - A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of 1 kb DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

otunon

Authorized by Derek Robinson 25 Sep 2017



Mulhiceheter

Inspected by Vanessa Mathieu-Sheltry 28 Nov 2017