240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: 1 kb DNA Ladder

Catalog #: N3232S/LConcentration: $500 \mu g/ml$ Unit Definition: N/ALot #: 1461712Assay Date: 12/2017Expiration Date: 12/2019Storage Temp: $-20^{\circ}C$

Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA

Specification Version: PS-N3232S/L v1.0
Effective Date: 30 Nov 2017

Assay Name/Specification (minimum release criteria)	Lot #1461712
A260/A280 Assay - The ratio of UV absorption of 1 kb DNA Ladder at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) - The concentration of 1 kb DNA Ladder is between 500 and 550 μ g/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern - The banding pattern of 1 kb DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each size fragment of the marker, when stained with Ethidium Bromide at a concentration of 0.5 μ g/ml.	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 μl reaction in 1X NEBuffer 2 containing 2.5 μg of 1 kb DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Authorized by Derek Robinson 30 Nov 2017







Inspected by

Vanessa Mathieu-Sheltry

15 Dec 2017