

New England Biolabs Certificate of Analysis

Product Name: *RsaI*
Catalog Number: *R0167S*
Concentration: *10,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Packaging Lot Number: *10102178*
Expiration Date: *03/2023*
Storage Temperature: *-20°C*
Storage Conditions: *100 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0167S/L v1.0*

| RsaI Component List | | | |
|---------------------|------------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| R0167SVIAL | RsaI | 10102177 | Pass |
| B7024AVIAL | Gel Loading Dye, Purple (6X) | 10091457 | Pass |
| B6004SVIAL | rCutSmart™ Buffer | 10103709 | Pass |

| Assay Name/Specification | Lot # 10102178 |
|--|----------------|
| Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 50 Units of RsaI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |
| Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 50 units of RsaI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity. | Pass |
| Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with RsaI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with RsaI. | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Pengda Zhang
Production Scientist
12 Apr 2021



Michael Tonello
Packaging Quality Control Inspector
12 Apr 2021