

New England Biolabs Certificate of Analysis

Product Name: Nspl
Catalog Number: R0602L
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10074547
Expiration Date: 05/2022
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200 µg/ml BSA
Specification Version: PS-R0602S/L v1.0

Nspl Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0602LVIAL	Nspl	10074546	Pass
B7204SVIAL	CutSmart® Buffer	10074630	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10064412	Pass

Assay Name/Specification	Lot # 10074547
Protein Purity Assay (SDS-PAGE) Nspl is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 50 Units of Nspl incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Nspl, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Nspl.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 10 units of Nspl incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass

Assay Name/Specification	Lot # 10074547
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 50 units of NspI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Penghua Zhang
Production Scientist
16 Jun 2020



Michael Tonello
Packaging Quality Control Inspector
16 Jun 2020