

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Nb.BssSI
Catalog Number: R0681S
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of pUC19 DNA in NEBuffer 3.1 incubated for 1 hour at 37°C in a

total reaction volume of 50 μl.

Packaging Lot Number: 10171511
Expiration Date: 12/2024
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 %

Glycerol , 500 μg/ml BSA, (pH 7.4 @ 25°C)

Specification Version: PS-R0681S v2.0

Nb.BssSI Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0681SVIAL	Nb.BssSI	10171510	Pass	
B6003SVIAL	NEBuffer™ r3.1	10168652	Pass	

Assay Name/Specification	Lot # 10171511
Protein Purity Assay (SDS-PAGE)	Pass
Nb.BssSl is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	
Non-Specific DNase Activity (16 hour)	Pass
A 50 µl reaction in NEBuffer 3.1 containing 1 µg of pUC19 DNA and a minimum of 20 units of Nb.BssSl incubated for 16 hours at 37°C results in a DNA pattern free of	
detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE:	
although no nuclease degradation is detected under these conditions, extended	
incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.	
Measured Activity (Restriction Endonuclease)	Pass
The measured activity of Nb.BssSI is complete at 20,000 units/ml and incomplete at 40,000 units/ml.	
Exonuclease Activity (Radioactivity Release)	Pass
A 50 μl reaction in NEBuffer 3.1 containing 1 μg of a mixture of single and	



R0681S / Lot: 10171511

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

YunJie Suń Production Scientist

08 Dec 2022

Michael Tonello

Packaging Quality Control Inspector

09 Dec 2022